

PCBs (as Aroclors *)

Offered by PT Provider (X = Yes)

		Catalog#	38043-EX	Cat. #832S	PEO-020	PE1033	PT-PCBW-WP
Analyte	CAS#	Absolute	ERA	NSI	RTC	Phenova	
PCBs as Aroclors		X	X	X	X	X	
1	Aroclor 1016 12674-11-2						
2	Aroclor 1221 11104-28-2						
3	Aroclor 1232 11141-16-5						
4	Aroclor 1242 53469-21-9						
5	Aroclor 1248 12672-29-6						
6	Aroclor 1254 11097-69-1						
7	Aroclor 1260 11096-82-5						

NOTE: PTs for PCBs are challenge samples. Only a single PCB is "spiked". The challenge is to identify and quantify the Aroclor present.

Accreditation for the PCB as Aroclors analyte group includes any other Aroclors not listed above.

A PT is **not** required for each individual Aroclor.

* PTs are currently not required to obtain/retain accreditation for the PCBs as Congeners analysis.

PT required for	
GC	GC/MS
✓	✓

s to correctly identify

ve.

lyte group.