



September 14, 2021

Kyle and Anna Bauer
6733 Crescent Ave.
Eau Claire Wi 54703

Dear Mr. & Ms. Bauer,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Bauer
DNR ID: 114

Page 21 of 26

Sample: 1273773 AH-11 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273773 AH-11 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	87%		1				S
1,2-Dichlorobenzene-d4 (SURR)	101%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Chris Bertsch
7593 Blue Valley Dr
Eau Claire WI 54703

Dear Mr. Bertsch:

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below the health advisory levels. Below is a list of chemicals detected:

Chemical	Result	Health Advisory
1,1-Dichloroethane	19	N/A
1,2-Dichloroethane	1.6	5
1,1-Dichloroethene	1.2	7
cis-1,2-Dichloroethene	7.6	70
1,2-Dichloropropane	0.69	5
Methylene chloride	1.1	5
Tetrachloroethene	2.3	5
1,1,1-Trichloroethane	1.6	200
1,1,2-Trichloroethane	0.86	5
Trichloroethene	0.78	5
Trichlorofluoromethane	0.36	N/A

If you have any questions regarding your results, please feel free to contact the writer at 715839-4734.

Sincerely,

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadaack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 1 of 26

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Bertsch

DNR ID: 234

Sample: 1273763 AH-1 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	19	ug/L	1	0.28	0.92		
1,2-Dichloroethane	1.6	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	1.2	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	7.6	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	[0.69]	ug/L	1	0.63	2.1	5	J
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	[1.1]	ug/L	1	1.1	3.7	5	J
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	2.3	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	1.6	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	[0.86]	ug/L	1	0.53	1.8	5	J
Trichloroethene	[0.78]	ug/L	1	0.46	1.5	5	J

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 2 of 26

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273763 - AH-1 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	[0.36]	ug/L	1	0.29	0.96		J
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	88%		1				S
1,2-Dichlorobenzene-d4 (SURR)	101%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

J = Result enclosed in brackets is between LOD and LOQ, a region of less certain quantitation.

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Kayla Briggs
7590 Crescent Ave
Eau Claire WI 54703

Dear Ms. Briggs:

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below the health advisory levels. Below is a list of chemicals detected:

Chemical	Result	Health Advisory
trans-1,2-Dichloroethene	0.45	100

If you have any questions regarding your results, please feel free to contact the writer at 715839-4734.

Sincerely,

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Briggs
DNR ID: 110

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273772-AH-10 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	[0.45]	ug/L	1	0.24	0.81	100	J
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273772 AH-10 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	83%		1				S
1,2-Dichlorobenzene-d4 (SURR)	95%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

J = Result enclosed in brackets is between LOD and LOQ, a region of less certain quantitation.

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Gerald Candell
7642 Blue Valley Dr.
Eau Claire Wi 54703

Dear Mr. Candell,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Page 17 of 26

Candell
DNR ID: 214

Sample: 1273771 AH-9 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 18 of 26

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273771 AH-9 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	84%		1				S
1,2-Dichlorobenzene-d4 (SURR)	99%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Lucas Cantu
6718 Crescent Ave.
Eau Claire Wi 54703

Dear Mr Cantu,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadaack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 23 of 26

Customer: Eau Claire County Health Department

NLS Project: 371960

Cantu
DNR ID: 224

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Sample: 1273774 AH-12 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273774 AH-12 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	87%		1				S
1,2-Dichlorobenzene-d4 (SURR)	101%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Jeff Freezy
7732 Blue Valley Dr.
Eau Claire Wi 54703

Dear Mr. Freezy,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Page 15 of 26

Freezy
DNR ID: 218

Sample: 1273770 AH-8 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273770 AH-8 Collected: 08/24/21 Analyzed: 09/07/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	86%		1				S
1,2-Dichlorobenzene-d4 (SURR)	99%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Terrance Godchaux
8720 Blue Valley Dr.
Eau Claire Wi 54703

Dear Mr. Godchaux,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Page 7 of 26

Godchaux
DNR ID: 226

Sample: 1273766 AH-4 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273766 AH-4 Collected: 08/24/21 Analyzed: 09/04/21 -- Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	92%		1				S
1,2-Dichlorobenzene-d4 (SURR)	101%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Tenille Goshaw
7617 Blue Valley Dr
Eau Claire WI 54703

Dear Ms. Goshaw:

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below the health advisory levels. Below is a list of chemicals detected:

Chemical	Result	Health Advisory
1,1-Dichloroethane	0.31	N/A

If you have any questions regarding your results, please feel free to contact the writer at 715839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Page 5 of 26

Goshaw
DNR ID: 204

Sample: 1273765 -AH-3 Collected: 08/24/21 Analyzed: 09/04/21 - Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	[0.31]	ug/L	1	0.28	0.92		J
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	



September 14, 2021

Jeff Reitz
7835 Blue Valley Dr.
Eau Claire Wi 54703

Dear Mr. Reitz,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadaack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 9 of 26

Customer: Eau Claire County Health Department

NLS Project: 371960

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Reitz
DNR ID: 232

Sample: 1273767 AH-5 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273767 AH-5 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	86%		1				S
1,2-Dichlorobenzene-d4 (SURR)	97%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.



September 14, 2021

Tyrel and Molly Zich
7865 Blue Valley Dr.
Eau Claire Wi 54703

Dear Mr. Godchaux,

Enclosed are the test results for volatile organic compounds (VOC's) for the water sample collected from your well on August 24th. All chemicals tested were below detection levels.

If you have any questions regarding your results, please feel free to contact the writer at 715-839-4734.

Sincerely,

Anna Hilger

Anna Hilger
Chemist, Eau Claire City-County Health Department

cc: Duabchi Vang, Wisconsin DNR
Dave Jadack, City of Eau Claire WWTP
Chuck Ahrens, GHD
Colleen Schian, City of Eau Claire

Enc

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 11 of 26

Customer: Eau Claire County Health Department NLS Project: 371960

Zich
DNR ID: 222

Project Description: VOC (Compliance)

Project Title:

Template: 524W Printed: 09/09/2021 12:02

Sample: 1273768 AH-6 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Benzene	ND	ug/L	1	0.43	1.4	5	
Bromobenzene	ND	ug/L	1	0.14	0.46		
Bromochloromethane	ND	ug/L	1	0.31	1.0		
Bromodichloromethane	ND	ug/L	1	0.42	1.4	80	
Bromoform	ND	ug/L	1	0.39	1.3	80	
Bromomethane	ND	ug/L	1	1.0	3.5		
n-Butylbenzene	ND	ug/L	1	0.49	1.6		
sec-Butylbenzene	ND	ug/L	1	0.41	1.4		
tert-Butylbenzene	ND	ug/L	1	0.51	1.7		
Carbon Tetrachloride	ND	ug/L	1	0.28	0.93	5	
Chlorobenzene	ND	ug/L	1	0.28	0.95	100	
Chloroethane	ND	ug/L	1	2.7	8.9		
Chloroform	ND	ug/L	1	0.52	1.7	80	
Chloromethane	ND	ug/L	1	0.40	1.3		
2-Chlorotoluene	ND	ug/L	1	0.36	1.2		
4-Chlorotoluene	ND	ug/L	1	0.40	1.3		
Dibromochloromethane	ND	ug/L	1	0.20	0.65	80	
1,2-Dibromo-3-Chloropropane	ND	ug/L	1	0.53	1.8		
1,2-Dibromoethane	ND	ug/L	1	0.28	0.95		
Dibromomethane	ND	ug/L	1	0.38	1.3		
1,2-Dichlorobenzene	ND	ug/L	1	0.12	0.40	600	
1,3-Dichlorobenzene	ND	ug/L	1	0.19	0.62		
1,4-Dichlorobenzene	ND	ug/L	1	0.22	0.73	75	
Dichlorodifluoromethane	ND	ug/L	1	0.35	1.2		
1,1-Dichloroethane	ND	ug/L	1	0.28	0.92		
1,2-Dichloroethane	ND	ug/L	1	0.43	1.4	5	
1,1-Dichloroethene	ND	ug/L	1	0.28	0.94	7	
cis-1,2-Dichloroethene	ND	ug/L	1	0.35	1.2	70	
trans-1,2-Dichloroethene	ND	ug/L	1	0.24	0.81	100	
1,2-Dichloropropane	ND	ug/L	1	0.63	2.1	5	
1,3-Dichloropropane	ND	ug/L	1	0.19	0.64		
2,2-Dichloropropane	ND	ug/L	1	0.87	2.9		
1,1-Dichloropropene	ND	ug/L	1	0.35	1.2		
cis-1,3-Dichloropropene	ND	ug/L	1	0.26	0.86		
trans-1,3-Dichloropropene	ND	ug/L	1	0.25	0.83		
Ethylbenzene	ND	ug/L	1	0.27	0.90	700	
Hexachlorobutadiene	ND	ug/L	1	0.60	2.0		
Isopropylbenzene	ND	ug/L	1	0.33	1.1		
p-Isopropyltoluene	ND	ug/L	1	0.46	1.5		
Methylene chloride	ND	ug/L	1	1.1	3.7	5	
Naphthalene	ND	ug/L	1	0.59	2.0		
n-Propylbenzene	ND	ug/L	1	0.40	1.3		
Styrene	ND	ug/L	1	0.31	1.0	100	
ortho-Xylene	ND	ug/L	1	0.28	0.95		
1,1,1,2-Tetrachloroethane	ND	ug/L	1	0.38	1.3		
1,1,1,2,2-Tetrachloroethane	ND	ug/L	1	0.60	2.0		
Tetrachloroethene	ND	ug/L	1	0.27	0.90	5	
Toluene	ND	ug/L	1	0.21	0.69	1000	
1,2,3-Trichlorobenzene	ND	ug/L	1	0.51	1.7		
1,2,4-Trichlorobenzene	ND	ug/L	1	0.44	1.5	70	
1,1,1-Trichloroethane	ND	ug/L	1	0.44	1.5	200	
1,1,2-Trichloroethane	ND	ug/L	1	0.53	1.8	5	
Trichloroethene	ND	ug/L	1	0.46	1.5	5	

ANALYTICAL RESULTS: GCMS 524.2, Rev 4.1 Safe Drinking Water Analysis

Page 12 of 26

Customer: Eau Claire County Health Department NLS Project: 371960

Project Description: VOC (Compliance)

Project Title: Template: 524W Printed: 09/09/2021 12:02

Sample: 1273768 AH-6 Collected: 08/24/21 Analyzed: 09/04/21 Analytes: 60

ANALYTE NAME	RESULT	UNITS	DIL	LOD	LOQ	MCL	Note
Trichlorofluoromethane	ND	ug/L	1	0.29	0.96		
1,2,3-Trichloropropane	ND	ug/L	1	0.44	1.5		
1,2,4-Trimethylbenzene	ND	ug/L	1	0.45	1.5		
1,3,5-Trimethylbenzene	ND	ug/L	1	0.43	1.4		
Vinyl chloride	ND	ug/L	1	0.19	0.62	.2	
meta,para-Xylene	ND	ug/L	1	0.59	2.0	10000	
MTBE	ND	ug/L	1	0.18	0.60		
4-Bromofluorobenzene (SURR)	87%		1				S
1,2-Dichlorobenzene-d4 (SURR)	98%		1				S

NOTES APPLICABLE TO THIS ANALYSIS:

S = This compound is a surrogate used to evaluate the quality control of a method.

VOLATILE ORGANIC CHEMICALS BY EPA METHOD 524.2

Eau Claire City-County Health Department DNR Lab Certification Number 618019820

OWNER:			Chris Bertsch	Chris Bertsch	Tenille Goshaw	Terrance Godchaux
SAMPLING DATE (YYMMDD):			210824	210824	210824	210824
LAB SAMPLE NUMBER:			210020	210021	210022	210023
	LOD	LOQ	Pre-treatment	Post-treatment		
CHEMICAL	ug/L	ug/L	ug/L	ug/L	ug/L	ug/L
1,1,1,2-tetrachloroethane	0.38	1.3	ND	ND	ND	ND
1,1,1-trichloroethane	0.44	1.5	1.6	ND	ND	ND
1,1,2,2-tetrachloroethane	0.60	2.0	ND	ND	ND	ND
1,1,2-trichloroethane	0.53	1.8	0.86	ND	ND	ND
1,1-dichloroethane	0.28	0.92	19	4.8	0.31	ND
1,1-dichloroethene	0.28	0.94	1.2	ND	ND	ND
1,1-dichloropropene	0.35	1.2	ND	ND	ND	ND
1,2,3-trichlorobenzene	0.51	1.7	ND	ND	ND	ND
1,2,3-trichloropropane	0.44	1.5	ND	ND	ND	ND
1,2,4-trichlorobenzene	0.44	1.5	ND	ND	ND	ND
1,2,4-trimethylbenzene	0.45	1.5	ND	ND	ND	ND
1,2-dibromo-3-chloropropane	0.53	1.8	ND	ND	ND	ND
1,2-dibromoethane	0.28	0.95	ND	ND	ND	ND
1,2-dichlorobenzene	0.12	0.40	ND	ND	ND	ND
1,2-dichloroethane	0.43	1.4	1.6	ND	ND	ND
1,2-dichloropropane	0.63	2.1	0.69	ND	ND	ND
1,3,5-trimethylbenzene	0.43	1.4	ND	ND	ND	ND
1,3-dichlorobenzene	0.19	0.62	ND	ND	ND	ND
1,3-dichloropropane	0.19	0.64	ND	ND	ND	ND
1,4-dichlorobenzene	0.22	0.73	ND	ND	ND	ND
2,2-dichloropropane	0.87	2.9	ND	ND	ND	ND
2-chlorotoluene	0.36	1.2	ND	ND	ND	ND
4-chlorotoluene	0.40	1.3	ND	ND	ND	ND
benzene	0.43	1.4	ND	ND	ND	ND
bromobenzene	0.14	0.46	ND	ND	ND	ND
bromochloromethane	0.31	1.0	ND	ND	ND	ND
bromodichloromethane	0.42	1.4	ND	ND	ND	ND
bromoform	0.39	1.3	ND	ND	ND	ND
bromomethane	1.0	3.5	ND	ND	ND	ND
carbon tetrachloride	0.28	0.93	ND	ND	ND	ND
chlorobenzene	0.28	0.95	ND	ND	ND	ND
chloroethane	2.7	8.9	ND	ND	ND	ND
chloroform	0.52	1.7	ND	ND	ND	ND
chloromethane	0.40	1.3	ND	ND	ND	ND
cis-1,2-dichloroethene	0.35	1.2	7.6	0.85	ND	ND
cis-1,3-dichloropropene	0.26	0.86	ND	ND	ND	ND
dibromochloromethane	0.20	0.65	ND	ND	ND	ND
dibromomethane	0.38	1.3	ND	ND	ND	ND
dichlorodifluoromethane	0.35	1.2	ND	ND	ND	ND
ethylbenzene	0.27	0.90	ND	ND	ND	ND
hexachlorobutadiene	0.60	2.0	ND	ND	ND	ND
isopropylbenzene	0.33	1.1	ND	ND	ND	ND
m,p-xylene	0.59	2.0	ND	ND	ND	ND
methylene chloride	1.1	3.7	1.1	ND	ND	ND
naphthalene	0.59	2.0	ND	ND	ND	ND
n-butylbenzene	0.49	1.6	ND	ND	ND	ND
o-xylene	0.28	0.95	ND	ND	ND	ND
para-isopropyltoluene	0.46	1.5	ND	ND	ND	ND
propylbenzene	0.4	1.3	ND	ND	ND	ND
sec-butylbenzene	0.41	1.4	ND	ND	ND	ND
styrene	0.31	1.0	ND	ND	ND	ND
t-butylbenzene	0.51	1.7	ND	ND	ND	ND
tetrachloroethylene	0.27	0.90	2.3	ND	ND	ND
toluene	0.21	0.69	ND	ND	ND	ND
trans-1,2-dichloroethene	0.24	0.81	ND	ND	ND	ND
trans-1,3-dichloropropene	0.25	0.83	ND	ND	ND	ND
trichloroethylene	0.46	1.5	0.78	ND	ND	ND
trichlorofluoromethane	0.29	0.96	0.36	ND	ND	ND
vinyl chloride	0.19	0.62	ND	ND	ND	ND

ND = Not detected

OWNER:		Jeff Reitz	Tyrel Zich	Sandy Bauer	Jeff Freezy
SAMPLING DATE (YYMMDD):		210824	210824	210824	210824
LAB SAMPLE NUMBER:		210024	210025	210026	210027
	LOD	LOQ			
CHEMICAL	ug/L	ug/L	ug/L	ug/L	ug/L
1,1,1,2-tetrachloroethane	0.38	1.3	ND	ND	ND
1,1,1-trichloroethane	0.44	1.5	ND	ND	ND
1,1,2,2-tetrachloroethane	0.60	2.0	ND	ND	ND
1,1,2-trichloroethane	0.53	1.8	ND	ND	ND
1,1-dichloroethane	0.28	0.92	ND	ND	ND
1,1-dichloroethene	0.28	0.94	ND	ND	ND
1,1-dichloropropene	0.35	1.2	ND	ND	ND
1,2,3-trichlorobenzene	0.51	1.7	ND	ND	ND
1,2,3-trichloropropane	0.44	1.5	ND	ND	ND
1,2,4-trichlorobenzene	0.44	1.5	ND	ND	ND
1,2,4-trimethylbenzene	0.45	1.5	ND	ND	ND
1,2-dibromo-3-chloropropane	0.53	1.8	ND	ND	ND
1,2-dibromoethane	0.28	0.95	ND	ND	ND
1,2-dichlorobenzene	0.12	0.40	ND	ND	ND
1,2-dichloroethane	0.43	1.4	ND	ND	ND
1,2-dichloropropane	0.63	2.1	ND	ND	ND
1,3,5-trimethylbenzene	0.43	1.4	ND	ND	ND
1,3-dichlorobenzene	0.19	0.62	ND	ND	ND
1,3-dichloropropane	0.19	0.64	ND	ND	ND
1,4-dichlorobenzene	0.22	0.73	ND	ND	ND
2,2-dichloropropane	0.87	2.9	ND	ND	ND
2-chlorotoluene	0.36	1.2	ND	ND	ND
4-chlorotoluene	0.40	1.3	ND	ND	ND
benzene	0.43	1.4	ND	ND	ND
bromobenzene	0.14	0.46	ND	ND	ND
bromochloromethane	0.31	1.0	ND	ND	ND
bromodichloromethane	0.42	1.4	ND	ND	ND
bromoform	0.39	1.3	ND	ND	ND
bromomethane	1.0	3.5	ND	ND	ND
carbon tetrachloride	0.28	0.93	ND	ND	ND
chlorobenzene	0.28	0.95	ND	ND	ND
chloroethane	2.7	8.9	ND	ND	ND
chloroform	0.52	1.7	ND	ND	ND
chloromethane	0.40	1.3	ND	ND	ND
cis-1,2-dichloroethene	0.35	1.2	ND	ND	ND
cis-1,3-dichloropropene	0.26	0.86	ND	ND	ND
dibromochloromethane	0.20	0.65	ND	ND	ND
dibromomethane	0.38	1.3	ND	ND	ND
dichlorodifluoromethane	0.35	1.2	ND	ND	ND
ethylbenzene	0.27	0.90	ND	ND	ND
hexachlorobutadiene	0.60	2.0	ND	ND	ND
isopropylbenzene	0.33	1.1	ND	ND	ND
m,p-xylene	0.59	2.0	ND	ND	ND
methylene chloride	1.1	3.7	ND	ND	ND
naphthalene	0.59	2.0	ND	ND	ND
n-butylbenzene	0.49	1.6	ND	ND	ND
o-xylene	0.28	0.95	ND	ND	ND
para-isopropyltoluene	0.46	1.5	ND	ND	ND
propylbenzene	0.4	1.3	ND	ND	ND
sec-butylbenzene	0.41	1.4	ND	ND	ND
styrene	0.31	1.0	ND	ND	ND
t-butylbenzene	0.51	1.7	ND	ND	ND
tetrachloroethylene	0.27	0.90	ND	ND	ND
toluene	0.21	0.69	ND	ND	ND
trans-1,2-dichloroethene	0.24	0.81	ND	ND	ND
trans-1,3-dichloropropene	0.25	0.83	ND	ND	ND
trichloroethylene	0.46	1.5	ND	ND	ND
trichlorofluoromethane	0.29	0.96	ND	ND	ND
vinyl chloride	0.19	0.62	ND	ND	ND

ND = Not detected

	OWNER:		Gerald Candell	Kayla Briggs	Kyle Bauer	Lucas Cantu
SAMPLING DATE (YYMMDD):			210824	210824	210824	210824
LAB SAMPLE NUMBER:			210028	210029	210030	210031
	LOD	LOQ				
CHEMICAL	ug/L	ug/L	ug/L	ug/L	ug/L	ug/L
1,1,1,2-tetrachloroethane	0.38	1.3	ND	ND	ND	ND
1,1,1-trichloroethane	0.44	1.5	ND	ND	ND	ND
1,1,2,2-tetrachloroethane	0.60	2.0	ND	ND	ND	ND
1,1,2-trichloroethane	0.53	1.8	ND	ND	ND	ND
1,1-dichloroethane	0.28	0.92	ND	ND	ND	ND
1,1-dichloroethene	0.28	0.94	ND	ND	ND	ND
1,1-dichloropropene	0.35	1.2	ND	ND	ND	ND
1,2,3-trichlorobenzene	0.51	1.7	ND	ND	ND	ND
1,2,3-trichloropropane	0.44	1.5	ND	ND	ND	ND
1,2,4-trichlorobenzene	0.44	1.5	ND	ND	ND	ND
1,2,4-trimethylbenzene	0.45	1.5	ND	ND	ND	ND
1,2-dibromo-3-chloropropane	0.53	1.8	ND	ND	ND	ND
1,2-dibromoethane	0.28	0.95	ND	ND	ND	ND
1,2-dichlorobenzene	0.12	0.40	ND	ND	ND	ND
1,2-dichloroethane	0.43	1.4	ND	ND	ND	ND
1,2-dichloropropane	0.63	2.1	ND	ND	ND	ND
1,3,5-trimethylbenzene	0.43	1.4	ND	ND	ND	ND
1,3-dichlorobenzene	0.19	0.62	ND	ND	ND	ND
1,3-dichloropropane	0.19	0.64	ND	ND	ND	ND
1,4-dichlorobenzene	0.22	0.73	ND	ND	ND	ND
2,2-dichloropropane	0.87	2.9	ND	ND	ND	ND
2-chlorotoluene	0.36	1.2	ND	ND	ND	ND
4-chlorotoluene	0.40	1.3	ND	ND	ND	ND
benzene	0.43	1.4	ND	ND	ND	ND
bromobenzene	0.14	0.46	ND	ND	ND	ND
bromochloromethane	0.31	1.0	ND	ND	ND	ND
bromodichloromethane	0.42	1.4	ND	ND	ND	ND
bromoform	0.39	1.3	ND	ND	ND	ND
bromomethane	1.0	3.5	ND	ND	ND	ND
carbon tetrachloride	0.28	0.93	ND	ND	ND	ND
chlorobenzene	0.28	0.95	ND	ND	ND	ND
chloroethane	2.7	8.9	ND	ND	ND	ND
chloroform	0.52	1.7	ND	ND	ND	ND
chloromethane	0.40	1.3	ND	ND	ND	ND
cis-1,2-dichloroethene	0.35	1.2	ND	ND	ND	ND
cis-1,3-dichloropropene	0.26	0.86	ND	ND	ND	ND
dibromochloromethane	0.20	0.65	ND	ND	ND	ND
dibromomethane	0.38	1.3	ND	ND	ND	ND
dichlorodifluoromethane	0.35	1.2	ND	ND	ND	ND
ethylbenzene	0.27	0.90	ND	ND	ND	ND
hexachlorobutadiene	0.60	2.0	ND	ND	ND	ND
isopropylbenzene	0.33	1.1	ND	ND	ND	ND
m,p-xylene	0.59	2.0	ND	ND	ND	ND
methylene chloride	1.1	3.7	ND	ND	ND	ND
naphthalene	0.59	2.0	ND	ND	ND	ND
n-butylbenzene	0.49	1.6	ND	ND	ND	ND
o-xylene	0.28	0.95	ND	ND	ND	ND
para-isopropyltoluene	0.46	1.5	ND	ND	ND	ND
propylbenzene	0.4	1.3	ND	ND	ND	ND
sec-butylbenzene	0.41	1.4	ND	ND	ND	ND
styrene	0.31	1.0	ND	ND	ND	ND
t-butylbenzene	0.51	1.7	ND	ND	ND	ND
tetrachloroethylene	0.27	0.90	ND	ND	ND	ND
toluene	0.21	0.69	ND	ND	ND	ND
trans-1,2-dichloroethene	0.24	0.81	ND	0.45	ND	ND
trans-1,3-dichloropropene	0.25	0.83	ND	ND	ND	ND
trichloroethylene	0.46	1.5	ND	ND	ND	ND
trichlorofluoromethane	0.29	0.96	ND	ND	ND	ND
vinyl chloride	0.19	0.62	ND	ND	ND	ND

ND = Not detected